



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

Scale No 1

DATE: May 10 2006 INITIALS: DEC OWNER: W/SR
DISTRICT: _____ ADDRESS: I-90 MP 300
LOCATION: Spokane POE No 64 COUNTY: Spokane
MAKE: Gold TYPE: truck scale ACCURACY CLASS: III
MODEL: 8530 ID. NO: 43437840V CAPACITY: 100,000 lbs SECTIONS: 2
MINIMUM GRADUATIONS: 20 lb REG. ELEMENT: digital PLATFORM: 12' X 20'

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
<u>Sec 1</u>	<u>20,000</u>	<u>19980</u>	<u>-20</u>				
<u>Center</u>	<u>20,000</u>	<u>20,000</u>	<u>0</u>				
<u>Sec 2</u>	<u>20,000</u>	<u>20,000</u>	<u>0</u>				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 2 # 2441-142

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David E Cronin Date: 8/23/2006



State of Washington
County of THURSTON
Signed and sworn to (or affirmed) before
me on 8-3-2006 by

Sylvia S. Lewallen
(Signature of notary public)
SYLVIA S. LEWALLEN
Printed Name of Notary
My appointment expires 3-28-2007

3000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS